



StarChill qPCR, Blue/White



Product Highlights

- Maintain sample temperature below 4°C for up to 2.5 hours*
- Innovative, color-changing material indicates when rack is approaching 7 °C
- Capacity: 64 x 0.1 ml Corbett Rotor-Gene® tubes, 16 x 0.2 ml single tubes (or 2 x 8-strips) and 4 x 0.5 ml tubes
- StarChill Rack must be frozen for a minimum of 8 hours at -20°C before use
- Freeze the StarChill Rack upside down in order for the gel to freeze evenly

Product Applications

Sample storage

STARLAB reserves the right to make changes at any time and without prior notice. The content and design of this PDF are protected by national and international copyright law and are the property of STARLAB International GmbH. Any duplication, editing, distribution and any kind of use and utilization of this PDF content in electronic systems, online media and / or libraries or similar databases requires the prior consent of STARLAB International GmbH.

Starlab (UK) LTD
5 Tanners Drive
MK14 5BU Milton Keynes
Email: infoline@starlab.co.uk



General Data

Art. No.	E2396-3008
Dimensions (W × D × H)	141 mm × 99 mm × 44 mm
Dimensions (W × D)	141 mm × 99 mm
Autoclavable	No
Pack Size	2 Pcs. (1 Pack × 2 Pcs.)
Rack Type	Chiller Rack

STARLAB reserves the right to make changes at any time and without prior notice. The content and design of this PDF are protected by national and international copyright law and are the property of STARLAB International GmbH. Any duplication, editing, distribution and any kind of use and utilization of this PDF content in electronic systems, online media and / or libraries or similar databases requires the prior consent of STARLAB International GmbH.

Starlab (UK) LTD
5 Tanners Drive
MK14 5BU Milton Keynes
Email: info@starlab.co.uk



More informations about StarChill qPCR, Blue/White

StarChill high performance chiller racks

Maintain a fridge environment on the benchtop! The centre section of StarChill racks changes colour when frozen. It then gradually turns back to its original colour as it warms up, thus giving an instant, visual-indication of the temperature.

*Time will vary, depending on ambient temperature. All StarChill racks must be frozen for a minimum of 8 hours at -20 °C before use. Not Autoclavable.